

Date: 16.09.2022

Tender Ref. No. GEM/2022/B/2500393

Item Name – Minus 80 Degree Deep Freezer

Technical Corrigendum-1

Sn. No.	Existing Tender Clause	Revised Tender clause
1	Technical Specification Point No.03 Should have a 5" (127mm) non-CFC foamed-in-place polyurethane insulation; 4.5" in door to provide best temperature uniformity.(documentary evidence to be submitted by the bidder)	Should have a minimum 125mm non-CFC foamed-in-place polyurethane insulation; minimum 4.5" in door to provide best temperature uniformity. (documentary evidence to be submitted by the bidder)
2	Technical Specification Point No.13 Should have two minimum10" tube axial fans to provide maximum cooling of the compressor housing	Should have minimum two fans to provide maximum cooling of the compressor housing