



UTTAR PRADESH MEDICAL SUPPLIES CORPORATION LTD.

Regd. Office: SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010

Ph: 0522 2838102 e-mail: drugs@upmsc.in

Provisional List of bidders/items of tender no. UPMSCL/Drugs-121/479 Date: 25.09.2021

In reference to below observations, if any firm have a clarification/representation regarding any firm/items, then it must be communicated to UPMSCL through e-mail i.e. drugs@upmsc.in before 17:00 PM, 27.09.2021. No communication thereafter in this regard shall be entertained by UPMSCL.

S. No	Name of Firm	Observation Related To Firm/Items	Provisional Status of the firm	Status of Sample
1	ALL TIME DATA PRIVATE LIMITED	<p>-Firm has not submitted all statutory notarized tender documents as given in requirements of cover A.</p> <p>-Firm has not submitted quality compliant with the below standards, or equivalent:</p> <ul style="list-style-type: none">- EU standard directive 93/42/EEC Class I, EN 455- EU standard directive 89/686/EEC Category III, EN 374- ANSI/SEA 105-2011- ASTM D6319-10 <p>-Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves.</p> <p>-Firm has not submitted average annual turnover statement with audited /Provisional balance sheet for the FY 2020-21.</p> <p>-Outer gloves preferably reach mid-forearm (minimum 280 mm total length)</p>	Not-Eligible	Submitted
2	Jajoo Surgical Pvt. Ltd.	<p>-Firm has not submitted all statutory notarized tender documents as given in requirements of cover A.</p> <p>Firm has not submitted average annual turnover statement with audited /Provisional balance sheet for the FY 2019-20, 2020-21.</p> <p>-Firm has not submitted quality compliant with the below standards, or equivalent:</p> <ul style="list-style-type: none">- EU standard directive 93/42/EEC Class I, EN 455- EU standard directive 89/686/EEC Category III, EN 374- ANSI/SEA 105-2011- ASTM D6319-10 <p>-Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves.</p> <p>-Firm has not submitted copy of authorization letter from manufacturer.</p> <p>-Outer gloves preferably reach mid-forearm (minimum 280 mm total length)</p>	Not-Eligible	Submitted
3	Kimberly Clark Hygiene Products Private Limited	<p>-Provisional balance sheet for the FY 2020-21 is submitted but not duly signed by auditor.</p> <p>-Firm has not submitted copy of valid ISO/Quality Management System Certificate/BIS/Any other quality certificates (as applicable for subject item as per annexure – A) issued by competent authority for all the premises.</p> <p>-Firm has not submitted quality compliant with the below standards, or equivalent:</p> <ul style="list-style-type: none">- EU standard directive 93/42/EEC Class I, EN 455- EU standard directive 89/686/EEC Category III, EN 374	Not-Eligible	Submitted

		<ul style="list-style-type: none"> - ANSI/SEA 105-2011 - ASTM D6319-10 -Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves. -Firm has not submitted all statutory notarized tender documents as given in requirements of cover A. -Firm has not submitted format XVII. -Firm has not submitted copy of authorization letter from manufacturer. -Outer gloves preferably reach mid-forearm (minimum 280 mm total length) 		
4	RFB LATEX LIMITED	<ul style="list-style-type: none"> -Firm has not submitted copy of valid license to manufacture/import the item(s) quoted as per specifications mentioned in the tender from the Competent Authority or DIC or FDA. -Firm has not submitted all statutory notarized tender documents as given in requirements of cover A. -Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves. -Test report is submitted by Satra tech which is located in UK but manufacturer is located in India. Kindly clarify. -Firm has not submitted quality compliant with the below standards, or equivalent: <ul style="list-style-type: none"> - EU standard directive 93/42/EEC Class I, EN 455 - EU standard directive 89/686/EEC Category III, EN 374 - ANSI/SEA 105-2011 - ASTM D6319-10 -Outer gloves preferably reach mid-forearm (minimum 280 mm total length) 	Not-Eligible	Submitted
5	Sanmati Packaging	<ul style="list-style-type: none"> -Firm has not submitted all statutory notarized tender documents as given in requirements of cover A. -Firm has not submitted average annual turnover statement with audited /Provisional balance sheet for the FY 2020-21. -Outer gloves preferably reach mid-forearm (minimum 280 mm total length) -Needs clarification regarding authenticity of test report submitted from SGS 	Not-Eligible	Submitted
6	Sky Lifecare Pvt Ltd	<ul style="list-style-type: none"> -Firm has not submitted format IV. -Firm has not submitted quality compliant with the below standards, or equivalent: <ul style="list-style-type: none"> - EU standard directive 93/42/EEC Class I, EN 455 - EU standard directive 89/686/EEC Category III, EN 374 - ANSI/SEA 105-2011 - ASTM D6319-10 -Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves. -Firm has not submitted all statutory notarized tender documents as given in requirements of cover A. -Firm has not submitted average annual turnover statement with audited /Provisional balance sheet for the FY 2020-21. -Firm has not submitted copy of authorization letter from manufacturer. -Outer gloves preferably reach mid-forearm (minimum 280 mm total length) 	Not-Eligible	Submitted
7	Sogo Fashions Private Limited	<ul style="list-style-type: none"> -Firm has not submitted all statutory notarized tender documents as given in requirements of cover A. -Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves. -Firm has not submitted average annual turnover statement with audited 	Not-Eligible	Submitted

		/Provisional balance sheet for the FY 2020-21. -Outer gloves preferably reach mid-forearm (minimum 280 mm total length)		
8	UNITY HEALTHCARE	-Firm has not submitted all statutory notarized tender documents as given in requirements of cover A. -Outer gloves preferably reach mid-forearm (minimum 280 mm total length) -Firm has not submitted complete test report from SGS.	Not-Eligible	Submitted

Managing Director
UPMSCL